



**PROPOSED EQUIPMENT LAYOUT**

**AGFA**   
**HealthCare**  
 DX-D 400  
 HIGH FREQUENCY  
 RADIOGRAPHIC SYSTEM SUITE

**EQUIPMENT SCHEDULE**

EQUIPMENT IDENTIFICATION NUMBER	DESCRIPTION	BTU / HOUR		WEIGHT (LBS.)	
		DIMENSIONS (IN.)		W	H
		W	D		
1	AGFA DX-D 400 SERIES HF GENERATOR CONTROL CONSOLE	17.00	11.70	4.80	18
2	AGFA DX-D 400 SERIES HF GENERATOR POWER CABINET	17.50	14.20	22.40	159
3	AGFA DX-D 400 ELEVATING 4-WAY FLOAT TOP TABLE	86.60	31.50	35.40	617
4	AGFA DX-D 400 FLOOR MOUNTED TUBESTAND	108.40	50.00	93.30	760
5	AGFA DX-D 400 VERTICAL RECEPTOR STAND	24.40	13.50	87.90	319
6	AGFA DX-D 400 DR SYSTEM WORKSTATION PC (WITH MONITOR, KEYBOARD AND MOUSE)	18.00	18.00	7.00	20
7	AGFA DX-D 400 DR SYSTEM UPS	13.00	6.00	8.00	10

 **GAMMA TECH, INC.**  
 1230 Creekview Ct. Voice (407) 556-3889  
 ST. Cloud, FL 34772 Cell (330) 360-4661  
 dhenry@gammatech.net www.gammatech.net

DESIGNER: DAVID HENRY

FILE NAME: DXD400.DWG  
 REVISION DATE: 09/20/2013  
 SCALE: TO FIT

- GENERAL NOTES**
- MINIMUM RECOMMENDED CEILING HEIGHT IS 8'-0".
  - SUGGESTED AMBIENT ROOM REQUIREMENTS ARE 60° TO 104° F. AT 30% TO 75% R.H. (NON-CONDENSING).
  - KW MODELS AVAILABLE 32KW, 50KW, 64KW, 80KW. MAIN POWER VARIES WITH KW MODEL.