



PROPOSED EQUIPMENT LAYOUT



**DX-D 500"
HIGH FREQUENCY
RADIOGRAPHIC SYSTEM SUITE**



GAMMA TECH, INC.

1213 Idaho Road
Austintown, OH 44515
dhenry@gammatech.net

Voice (330) 792-9568
Fax (330) 792-0658
www.gammatech.net

DESIGNER: DAVID HENRY

FILE NAME: DXD500N.DWG
REVISION DATE: 01/26/2010
SCALE: TO FIT

EQUIPMENT SCHEDULE

EQUIPMENT IDENTIFICATION NUMBER

		BTU / HOUR	WEIGHT (LBS.)				
			DESCRIPTION	DIMENSIONS (IN.)			
				W	x	D	x
1	500	AGFA QG-80 80KW GENERATOR GENERATOR AND DR WORKSTATION PC	TYPICAL PC				
2	3600	AGFA QG-80 80KW GENERATOR DEDICATED GENERATOR CABINET	24.00	20.00	33.50	380	
3	400	AGFA QT-750 ELEVATING 4-WAY FLOAT TOP TABLE	85.00	35.55	32.50	500	
4	1200	AGFA RS-580 CEILING SUSPENDED X-RAY TUBE CRANE	188.00	118.50	112.79	1200	
5	35	AGFA QW-420 VERTICAL RECEPTOR STAND	26.00	14.50	84.00	225	

GENERAL NOTES

- MINIMUM RECOMMENDED CEILING HEIGHT IS 9'-0".
- SUGGESTED AMBIENT ROOM REQUIREMENTS ARE 66° TO 75° F. AT 30% TO 60% R.M. (NON-CONDENSING).
- MAIN INCOMING POWER SUPPLY: 480VAC, ±5%, THREE PHASE, 100KVA.