



GENERAL NOTES

- ALL SPECIFICATIONS AND SPACE REQUIREMENTS ARE SHOWN GENERICALLY. CONTACT GAMMA TECHNOLOGIES FOR SPECIFIC REQUIREMENTS PERTAINING TO OPTIONAL EQUIPMENT COMPONENTS AND ADDITIONAL ROOM PREPARATION SPECIFICATIONS.
- MINIMUM RECOMMENDED CEILING HEIGHT IS 8'-0"
- SUGGESTED AMBIENT ROOM REQUIREMENTS ARE 20-25° C [68-77° F] AT 40% TO 60% NON-CONDENSING.
- MAIN INCOMING POWER SUPPLY:
208VAC, THREE PHASE, 20AMP POWER SOURCE

EQUIPMENT SCHEDULE

		EQUIPMENT IDENTIFICATION NUMBER				
		BTU / HOUR	WEIGHT (LBS.)			
		DESCRIPTION	DIMENSIONS (IN.) W x D x H			
1	6500	GENERAL ELECTRIC MEDICAL SYSTEMS MILLENNIUM VG IMAGING SYSTEM GANTRY	61.00	69.00	82.70	5842
2		GENERAL ELECTRIC MEDICAL SYSTEMS MILLENNIUM VG IMAGING SYSTEM IMAGING TABLE	27.50	99.00	????	882
3		GENERAL ELECTRIC MEDICAL SYSTEMS MILLENNIUM VG IMAGING SYSTEM COLLIMATOR CART	34.20	19.30	49.00	992
4	1700	GENERAL ELECTRIC MEDICAL SYSTEMS MILLENNIUM VG IMAGING SYSTEM ACQUISITION COMPUTER	8.70	26.00	20.50	66
5	342	GENERAL ELECTRIC MEDICAL SYSTEMS MILLENNIUM VG IMAGING SYSTEM ACQUISITION MONITOR/KEYBOARD	17.00	18.00	19.30	82
6		GENERAL ELECTRIC MEDICAL SYSTEMS MILLENNIUM VG IMAGING SYSTEM eNTEGRA WORKSTATION (OPTIONAL)	????	????	????	????



GAMMA TECH, INC
 129 Wild Berry Lane
 Hampstead, NC 28443
 Ph (910) 270-9787
 Fax (910) 270-9767
 www.gammatech.net

© PLANNING DIV.
 FILE NAME:
 GE MILLENNIUM VG NUC-MED
 REVISION DATE:
 01.31.2008