


TYPICAL ROOM AND EQUIPMENT LAYOUT
 (DASHED LINES DEPICT TABLE TOP TRAVEL AND SERVICE CLEARANCES)



GENERAL NOTES

- ALL SPECIFICATIONS AND SPACE REQUIREMENTS ARE SHOWN GENERICALLY. CONTACT GAMMA TECH, INC. FOR SPECIFIC REQUIREMENTS PERTAINING TO OPTIONAL EQUIPMENT COMPONENTS AND ADDITIONAL ROOM PREPARATION SPECIFICATIONS.
- MINIMUM RECOMMENDED CEILING HEIGHT 9'-0"
- SUGGESTED AMBIENT ROOM REQUIREMENTS: 55 - 75° F
 RELATIVE HUMIDITY: 10% - 75% NON-CONDENSING
- MAIN INCOMING POWER SUPPLY :
 480VAC, THREE PHASE, 150KVA, DELTA OR WYE,
 80 AMP, 3-POLE, MAIN CB

EQUIPMENT SCHEDULE

EQUIPMENT IDENTIFICATION NUMBER

		BTU / HOUR	WEIGHT (LBS.)				
			DESCRIPTION	DIMENSIONS (IN.)			
				W	x	D	x
1	204	GENERAL ELECTRIC MEDICAL SYSTEMS PROTEUS XR/a SYSTEM OPERATORS CONSOLE (ON PEDESTAL)	13.78	9.06	2.36	6	
2	2037	GENERAL ELECTRIC MEDICAL SYSTEMS PROTEUS XR/a SYSTEM GENERATOR CABINET	21.65	17.72	63.00	331	
3	1706	GENERAL ELECTRIC MEDICAL SYSTEMS PROTEUS XR/a SYSTEM TABLE	90.33	34.65	31.50	441	
4	1200	GENERAL ELECTRIC MEDICAL SYSTEMS XT RADIOGRAPHIC SUSPENSION WITH INBOARD MOUNTINGS	228.00	120.50	94.00 MODIFIED TO FIT ROOM	941	
5	150	GENERAL ELECTRIC MEDICAL SYSTEMS PROTEUS XR/a SYSTEM WALL STAND	24.00	15.00	91.28	221	

**PROTEUS XR/a SYSTEM
 RADIOGRAPHIC SUITE**

GAMMA TECH, INC



129 Wild Berry Lane
 Hampstead, NC 28443
 Ph (910) 270-9787
 Fax (910) 270-9767
 www.gammatech.net

© PLANNING DIVISION

FILE NAME:
 GE PROTEUS XRa

REVISION DATE:
 07.16.2008