


TYPICAL ROOM AND EQUIPMENT LAYOUT



GENERAL NOTES

- ALL SPECIFICATIONS AND SPACE REQUIREMENTS ARE SHOWN GENERICALLY. CONTACT GAMMA TECHNOLOGIES FOR SPECIFIC REQUIREMENTS PERTAINING TO OPTIONAL EQUIPMENT COMPONENTS AND ADDITIONAL ROOM PREPARATION SPECIFICATIONS.
- RANGE OF LINE VOLTAGES :
NOMINAL LINE VOLTAGE OF 200 TO 240 1 PHASE, 50 OR 60 Hz.
- AMBIENT OPERATION TEMPERATURE: 50-104° F (10-40° C)
ALLOWABLE TEMPERATURE CHANGE: 15° F (10° C) per hour
HUMIDITY: 30-75 PERCENT NON-CONDENSING

EQUIPMENT SCHEDULE

	EQUIPMENT IDENTIFICATION NUMBER	BTU / HOUR		WEIGHT (LBS.)			
		DESCRIPTION		DIMENSIONS (IN.)			
				W	x	D	x
1	4952	GE SENOGRAPHE 800T W/ INTEGRATED SHIELD	46.00	51.00	????	666	
2	976	GE SENOGRAPHE 800T EDM TOWER (OPTIONAL)	6.75	16.75	????	33	

GAMMA TECH, INC



129 Wild Berry Lane
Hampstead, NC 28443
Ph (910) 270-9787
Fax (910) 270-9767
www.gammatech.net

© PLANNING DIVISION

FILE NAME:
GE SENOGAPHE 800T

REVISION DATE:
01.31.2008



GE Medical Systems

SENOGRAPHE 800T
MAMMOGRAPHY SYSTEM