

TYPICAL ROOM AND EQUIPMENT LAYOUT

(DASHED LINES DEPICT EQUIPMENT TRAVEL AND SERVICE CLEARANCES)
RECOMMENDED FINISHED CEILING HEIGHT = 8'-0"



www.gammatech.net



GE Medical Systems

GE SYTEC 2/3/4000I CT SCANNER

GENERAL NOTES

- ALL SPECIFICATIONS AND SPACE REQUIREMENTS ARE SHOWN GENERICALLY.
 CONTACT GAMMA TECH, INC. FOR SPECIFIC REQUIREMENTS PERTAINING TO
 OPTIONAL EQUIPMENT COMPONENTS AND ADDITIONAL ROOM PREPARATION SPECIFICATIONS.
- RECOMMENDED FINISHED CEILING HEIGHT IS 8'-0".
 MINIMUM FINISHED CEILING HEIGHT 7'-6".
- AMBIENT OPERATION TEMPERATURE: 60° TO 75° F. HUMIDITY: 30% TO 60% (NON-CONDENSING).
 ALLOWABLE TEMPERATURE CHANGE: 5°F PER HOUR ALLOWABLE HUMIDITY CHANGE: 10% PER HOUR

MAX - ALTITUDE: 10,000FT ABOVE SEA LEVEL

© PLANNING DIVISION

FILE NAME: GE SYTEC 2_3_4000I

REVISION DATE: 01.23.2008

EQUIPMENT SCHEDULE

EQUIPMENT IDENTIFICATION NUMBER

	BTU / HOUR		WEIGHT (LBS.)	
		DESCRIPTION	DIMENSIONS (IN.)	
1	5955	GE MEDICAL SYSTEMS CT SYTEC 2/3/4000I GANTRY		1716
2		GE MEDICAL SYSTEMS PATIENT TABLE W/ EXTENDED TABLE		800
3	1985	GE MEDICAL SYSTEMS I SERIES OPERATOR'S CONSOLE		220
4	794	GE MEDICAL SYSTEMS X-RAY GENERATOR (MUST BE LOCATED A MIN. OF 6'-0" GANTRY AND/OR CONSOLE)		755
5		GE MEDICAL SYSTEMS REMOTE CONTROL FOR INJECTOR		4
6	321	GE MEDICAL SYSTEMS INJECTOR ELECTRONICS		37
7		GE MEDICAL SYSTEMS INJECTOR HEAD ON OVERHEAD COUNTERPOISED SUSPENSION		79
		PowerTECH SUPPRESSOR/FILTER WALL MOUNTED		