



**TYPICAL ROOM AND EQUIPMENT LAYOUT**  
 (DASHED LINES DEPICT TABLE TOP TRAVEL AND SFD CLEARANCE.)

### GENERAL NOTES

- ALL SPECIFICATIONS AND SPACE REQUIREMENTS ARE SHOWN GENERICALLY. CONTACT GAMMA TECHNOLOGIES FOR SPECIFIC REQUIREMENTS PERTAINING TO OPTIONAL EQUIPMENT COMPONENTS AND ADDITIONAL ROOM PREPARATION SPECIFICATIONS.
- MINIMUM RECOMMENDED CEILING HEIGHT IS 8'-0"
- SUGGESTED AMBIENT ROOM REQUIREMENTS ARE 59-81° F [15-27° C] AT 20% TO 80% NON-CONDENSING.
- MAIN INCOMING POWER SUPPLY:  
120VAC, SINGLE PHASE, 15AMP POWER SOURCE

### EQUIPMENT SCHEDULE

		EQUIPMENT IDENTIFICATION NUMBER				
		BTU / HOUR	WEIGHT (LBS.)			
		DESCRIPTION	DIMENSIONS (IN.)			
			W	x	D	x
1	1100	GENERAL ELECTRIC MEDICAL SYSTEMS 600XR TOMOGRAPHIC SYSTEM	99.40	198.00	73.00	2559
2		GENERAL ELECTRIC MEDICAL SYSTEMS 600XR SIZE COLLIMATOR STACKER	33.50	23.60	47.40	1003
3		GENERAL ELECTRIC MEDICAL SYSTEMS 600XR SIZE COLLIMATOR CART	27.20	34.60	51.20	406
4		GENERAL ELECTRIC MEDICAL SYSTEMS UNIVERSAL CARBON FIBER IMAGING TABLE ON STANDARD SCANNING RAILS	28.00	172.25	35.25	496
5	4500	GENERAL ELECTRIC MEDICAL SYSTEMS STARCAM 4000I ELECTRONICS CABINET	18.07	33.20	25.60	220
6	215	GENERAL ELECTRIC MEDICAL SYSTEMS ALPHA/NUMERIC MONITOR AND KEYBOARD	26.00	10.40	2.00	55
7	215	GENERAL ELECTRIC MEDICAL SYSTEMS HIGH RESOLUTION COLOR MONITOR	29.00	????	18.00	55
8	510	GENERAL ELECTRIC MEDICAL SYSTEMS MULTIFORMATTER HD VIDEO (OPTIONAL)	15.19	17.63	82.28	95



**GE Medical Systems**  
 XRT 4000I - NUCLEAR MEDICINE SUITE

**GAMMA TECH, INC**



129 Wild Berry Lane  
 Hampstead, NC 28443  
 Ph (910) 270-9787  
 Fax (910) 270-9767  
 www.gammatech.net

© PLANNING DIV.

FILE NAME:  
 GE XRT 4000I NUC-MED  
 REVISION DATE:  
 01.31.2008