



SCALE

TYPICAL ROOM AND EQUIPMENT LAYOUT



## EQUIPMENT SCHEDULE

### GENERAL NOTES

- ALL SPECIFICATIONS AND SPACE REQUIREMENTS ARE SHOWN GENERICALLY. CONTACT GAMMA TECH, INC. FOR SPECIFIC REQUIREMENTS PERTAINING TO OPTIONAL EQUIPMENT COMPONENTS AND ADDITIONAL ROOM PREPARATION SPECIFICATIONS.
- MAIN INCOMING POWER SUPPLY: INPUT LINE <A>: 480 VAC, 100 AMP THREE PHASE
- MAIN INCOMING POWER SUPPLY: INPUT LINE <A1>: 208 VAC, 30 AMP SINGLE PHASE

### EQUIPMENT IDENTIFICATION NUMBER

ID	BTU / HOUR	DESCRIPTION	WEIGHT (LBS.)			
			DIMENSIONS (IN.)			
			W	x	D	x
1	132	LODOX STATSCAN DIGITAL X-RAY CONTROL CONSOLE	31.50	19.69	59.05	88
2	1200	LODOX STATSCAN DIGITAL X-RAY GENERATOR	24.80	17.72	22.05	133
3	2668	LODOX STATSCAN DIGITAL X-RAY SCANNER	114.17	98.42	118.11	2668
4	N/A	LODOX STATSCAN TROLLEY	29.13	86.61	31.89	177



© PLANNING DIVISION

FILE NAME:  
LODOX STATSCAN

REVISION DATE:  
07.17.2008