


TYPICAL ROOM AND EQUIPMENT LAYOUT



GENERAL NOTES

- ALL SPECIFICATIONS AND SPACE REQUIREMENTS ARE SHOWN GENERICALLY. CONTACT GAMMA TECH, INC. FOR SPECIFIC REQUIREMENTS PERTAINING TO OPTIONAL EQUIPMENT COMPONENTS AND ADDITIONAL ROOM PREPARATION SPECIFICATIONS.
- INPUT LINE: 200-240 VAC (±10%), 25 AMP, SINGLE PHASE 50-60HZ
- VERTICAL TRAVEL: 28-55 IN (71-140CM) (MOTORIZED)
ROTATION: +195° TO -150°
SID: 65 CM
PATIENT FACE SHIELD: REMOVABLE

EQUIPMENT SCHEDULE

EQUIPMENT IDENTIFICATION NUMBER		BTU / HOUR	WEIGHT (LBS.)		
		DESCRIPTION	DIMENSIONS (IN.)		
			W	D	H
1	1200	LORAD AFFINITY MAMMOGRAPHY	55.00	43.00	89.00
					589

GAMMA TECH, INC



129 Wild Berry Lane
Hampstead, NC 28443
Ph (910) 270-9787
Fax (910) 270-9767
www.gammatech.net

© PLANNING DIVISION

FILE NAME:
LORAD AFFINITY

REVISION DATE:
07.17.2008

LORAD

AFFINITY
MAMMOGRAPHY SYSTEM