



# TYPICAL ROOM LAYOUT



## GENERAL NOTES

- ALL SPECIFICATIONS AND SPACE REQUIREMENTS ARE SHOWN GENERICALLY. CONTACT GAMMA TECH, INC. FOR SPECIFIC REQUIREMENTS PERTAINING TO OPTIONAL EQUIPMENT COMPONENTS AND ADDITIONAL ROOM PREPARATION SPECIFICATIONS.
- RECOMMENDED CEILING HEIGHT IS 8'-0".
- SUGGESTED AMBIENT ROOM REQUIREMENTS ARE 66° TO 75° F. AT 30% TO 60% R.M. (NON-CONDENSING).
- MAIN INCOMING POWER SUPPLY: 220VAC, SINGLE PHASE, 20AMP, 6-20R

## EQUIPMENT SCHEDULE

		EQUIPMENT IDENTIFICATION NUMBER				
		BTU / HOUR	WEIGHT (LBS.)			
		DESCRIPTION	DIMENSIONS (IN.)			
			W	x	D	x
1	2100	LORAD M-IIIE MAMMOGRAPHY SYSTEM	41.00	58.00	77.00	600

## LORAD M-IIIE MAMMOGRAPHY SYSTEM SUITE

**GAMMA TECH, INC**



129 Wild Berry Lane  
Hampstead, NC 28443  
Ph (910) 270-9787  
Fax (910) 270-9767  
www.gammatech.net

© PLANNING DIVISION

FILE NAME:  
LORAD MIIIE

REVISION DATE:  
07.17.2008