

EQUIPMENT SCHEDULE						
EQUIPMENT IDENTIFICATION NUMBER						
	BTU / HOUR	DESCRIPTION	WEIGHT (LBS.)			
			W	D		
			x	x		
			H			
1	3790	SIEMENS SOMATOM PLUS GANTRY	105.00	39.30	83.00	5940
2	230	SIEMENS SOMATOM PLUS PATIENT TABLE	28.00	83.90	26.80	990
3	886	SIEMENS SOMATOM PLUS CONTROL CONSOLE	47.20	30.70	29.50	330
	300	SIEMENS SOMATOM PLUS CONTROL CONSOLE MONITORS	14.50	14.20	13.00	26
4	5464	SIEMENS SOMATOM PLUS X-RAY GENERATOR CABINET	25.90	25.60	83.50	869
5	4360	SIEMENS SOMATOM PLUS POWER DISTRIBUTION CABINET	23.60	33.70	84.00	1366
6	8196	SIEMENS SOMATOM PLUS COMPUTER/IMAGE PROCESSOR	23.60	33.70	84.00	825
7	342	SIEMENS SOMATOM PLUS HEAT EXCHANGER C	25.20	29.50	31.50	211
8	N/A	MEDRAD MARK IV INJECTOR HEAD ON MOBILE STAND	26.00	26.00	60.25	30
9	320	MEDRAD MARK IV ELECTRONICS CABINET	19.00	12.59	10.50	43


EQUIPMENT SCHEDULE						
EQUIPMENT IDENTIFICATION NUMBER						
	BTU / HOUR	DESCRIPTION	WEIGHT (LBS.)			
			W	D		
			x	x		
			H			
10	N/A	MEDRAD MARK IV REMOTE PANEL (ON COUNTER)	16.53	3.70	5.38	5
11	500	SONY FILM STATION DRY FILM IMAGER	23.60	27.13	9.88	122
12	1500	SIEMENS UPS MOUNTED UNDER COUNTER				
13		FILTRINE WATER CHILLER MOUNTED ON ROOF				

SIEMENS Somatom Plus-4

MEDICAL SYSTEMS, INC. CAT Scanner suite

● ALL SPECIFICATIONS AND SPACE REQUIREMENTS ARE SHOWN GENERICALLY. CONTACT GAMMA TECHNOLOGIES FOR SPECIFIC REQUIREMENTS PERTAINING TO OPTIONAL EQUIPMENT COMPONENTS AND ADDITIONAL ROOM PREPARATION SPECIFICATIONS.

● MAIN INCOMING POWER SUPPLY: INPUT LINE: 480VAC, 3-PHASE, 150AMP



GAMMA TECH, INC
 129 Wild Berry Lane Ph (910) 270-9787
 Hampstead, NC 28443 Fax (910) 270-9767
www.gammatech.net

© PLANNING DIVISION
 FILE NAME: SOMATOM PLUS
 REV. DATE: 08.31.2011