



TYPICAL ROOM AND EQUIPMENT LAYOUT

SIEMENS DUAL HEAD E.CAM
MEDICAL SYSTEMS, INC. NUCLEAR MEDICINE
 SYSTEM SUITE

GENERAL NOTES

- ALL SPECIFICATIONS AND SPACE REQUIREMENTS ARE SHOWN GENERICALLY. CONTACT GAMMA TECHNOLOGIES FOR SPECIFIC REQUIREMENTS PERTAINING TO OPTIONAL EQUIPMENT COMPONENTS AND ADDITIONAL ROOM PREPARATION SPECIFICATIONS.
- MAIN INCOMING POWER SUPPLY:
 INPUT LINE: 208VAC, 60 AMP
 SINGLE PHASE

TEMPERATURE AND OPERATION:

- OPERATING TEMPERATURE SHALL BE 60° to 95°F. TO AVOID PERMANENT DAMAGE TO THE SCINTILLATION CRYSTAL, TEMPERATURE VARIANCE DURING OPERATION OR STORAGE MUST NOT EXCEED 8°F PER HOUR.

HUMIDITY:

- BETWEEN 15% AND 80% NON-CONDENSING.

EQUIPMENT SCHEDULE

EQUIPMENT IDENTIFICATION NUMBER	DESCRIPTION	BTU / HOUR		WEIGHT (LBS.)	
		DIMENSIONS (IN.)			
		W	x	D	x
1	SIEMENS E.CAM GANTRY W/ DUAL DETECTORS	66.00	62.75	76.00	4,239
2	SIEMENS E.CAM PATIENT HANDLING SYSTEM (TABLE)	35.43	99.00	43.00	562
3	5850 SIEMENS E.CAM ELECTRONICS TOWER	22.00	17.75	41.50	220
4	950 SIEMENS E.SOFT ACQUISITION CPU (UNDER TABLE)	8.00	16.68	19.00	31
5	409 SIEMENS E.SOFT ACQUISITION MONITOR (ON TABLE)	20.25	19.50	21.00	55
6	81 SIEMENS E.SOFT READ/WRITE DRIVE, CD-R (ON TABLE)	7.13	12.13	2.13	2
7	SIEMENS E.CAM COLLIMATOR CART (EMPTY)	31.75	34.50	52.00	256
8	950 SIEMENS E.SOFT VIEWING STATION CPU	10.00	17.50	18.00	31
9	409 SIEMENS E.SOFT VIEWING STATION MONITOR	20.25	19.50	21.00	55
10	ALPHA 2500 208V IN/OUT UPS (UNINTERRUPTABLE POWER SUPPLY)	8.50	22.00	20.50	

GAMMA TECH, INC



129 Wild Berry Lane
 Hampstead, NC 28443
 Ph (910) 270-9787
 Fax (910) 270-9767
www.gammatech.net

© PLANNING DIVISION

FILE NAME:
 TYP ECAM

REVISION DATE:
 09.01.2011