



# EQUIPMENT SCHEDULE

NOTES	EQUIPMENT IDENTIFICATION NUMBER						
		BTU / HOUR	WEIGHT (LBS.)				
			DESCRIPTION	DIMENSIONS (IN.)			
			W	x	D	x	H
OFOI	1	275	SIEMENS OPEN VIVA M.R.I. MR OPERATING KEYBOARD	27.00	10.00	3.00	10
OFOI	2	N/A	SIEMENS OPEN VIVA M.R.I. CONTROL MONITOR	19.00	20.00	18.75	45
OFOI	3	N/A	SIEMENS OPEN VIVA M.R.I. ALARM BOX / MAGNET STOP	12.00	4.00	18.75	3
OFOI	4	3,415	SIEMENS OPEN VIVA M.R.I. 0.2T VIVA MAGNET WITH COVERS	48.00	79.00	77.00	25360
OFOI	5	N/A	SIEMENS OPEN VIVA M.R.I. PATIENT TABLE	31.13	83.00	33.00	440
OFOI	6	515	SIEMENS OPEN VIVA M.R.I. RF FILTER PLATE WITH RFCE	34.50	19.50	46.50	176
OFOI	7	N/A	SIEMENS OPEN VIVA M.R.I. MAGNET STOP	3.00	5.00	3.00	1
OFOI	8	N/A	SIEMENS OPEN VIVA M.R.I. INTERCOM SPEAKER	4.00	3.00	11.00	4
OFOI	9	27,325*	SIEMENS OPEN VIVA M.R.I. CONTROL CABINET	74.00	34.00	71.00	3310
OFOI	10		SIEMENS OPEN VIVA M.R.I. POWER CABINET				
OFOI	11		SIEMENS OPEN VIVA M.R.I. GRADIENT POWER CABINET				
OFOI	12	155	SIEMENS OPEN VIVA M.R.I. HALL PROBE (OPTION)	5.13	8.68	13.38	3
OFOI	13	52	SIEMENS OPEN VIVA M.R.I. EFI-2 UNIT (OPTION)	20.00	21.00	15.00	44
OFOI	14	1,327	SIEMENS OPEN VIVA M.R.I. WDC TECHNOLOGIES U.P.S.	60.00	33.00	71.00	3550
OFOI	15	N/A	SIEMENS OPEN VIVA M.R.I. (REMOTELY LOCATED) WATER CHILLER	40.00	68.00	73.00	1525
OFOI	16	N/A	SIEMENS OPEN VIVA M.R.I. FLOW INDICATOR PANEL	24.00	6.00	34.00	

NOTE EXPLANATION:  
 CFCI = CONTRACTOR FURNISHED - CONTRACTOR INSTALLED  
 OFCI = OWNER FURNISHED - CONTRACTOR INSTALLED  
 OFOI = OWNER FURNISHED - OWNER INSTALLED  
 \* GPS THREE CABINETS THAT MAY NOT BE SEPARATED.

PLANNING BY:



**GAMMA TECH, INC**  
 129 Wild Berry Lane  
 Hampstead, NC 28443  
 Ph (910) 270-9787 • Fax (910) 270-9767  
 www.gammatech.net

**SIEMENS**  
**MEDICAL SYSTEMS, INC.**  
**OPEN VIVA M.R.I. SUITE**

**GENERAL NOTES**

- ALL SPECIFICATIONS AND SPACE REQUIREMENTS ARE SHOWN GENERICALLY. CONTACT GAMMA TECHNOLOGIES FOR SPECIFIC REQUIREMENTS PERTAINING TO OPTIONAL EQUIPMENT COMPONENTS AND ADDITIONAL ROOM PREPARATION SPECIFICATIONS.

© PLANNING DIVISION  
 FILE NAME:  
 TYP OPEN VIVA  
 REVISION DATE: 09.01.2011